

برگشت (/IACLD/Panel/Labs/Answers/29630)

دوره چهل و دوم تابستان 1401

S2 Results

نمونه مجهول ToRCH

کد نمونه 1161	EQAP-18136	آزمایشگاه پاتوبیولوژی و ژنتیک رصد
---------------	------------	-----------------------------------

Anti-CMV (IgG)	Anti-HSV1 (Anti-Herpes Type1 IgG)	Anti-Rubella (IgG)	Anti-Toxoplasma (IgG)
----------------	-----------------------------------	--------------------	-----------------------

Anti-CMV (IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان بر اساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است.

- ثبت نام کنندگان 215
- شرکت کنندگان 195
- کل درصد مشارکت 90.70 %
- پاسخهای دریافتی 170
- درصد پاسخدهی 87.18 %

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	61	59	0	2	%96.72	%0.00	%3.28
ECL	22	22	0	0	%100.00	%0.00	%0.00
ELFA	17	17	0	0	%100.00	%0.00	%0.00
ELISA	68	64	0	4	%94.12	%0.00	%5.88
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00
Other	2	2	0	0	%100.00	%0.00	%0.00

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
CL	Abbott	Positive	11	
CL	Architect	Positive	3	
CL	Autobio Diagnostic	Positive	4	
CL	Biomerieux	Positive	1	

Method	Kit	Result	Count	Your Answer
CL	Dia Sorin	Negative	2	
CL	Dia Sorin	Positive	23	←
CL	Diaplus	Positive	1	
CL	Liaison	Positive	5	
CL	Siemens	Positive	10	
ECL	Abbott	Positive	1	
ECL	Architect	Positive	1	
ECL	Biomerieux	Positive	1	
ECL	Cobas	Not Reported	1	
ECL	Roche	Positive	19	
ELFA	Biomerieux	Positive	16	
ELFA	Diesse-Tec	Positive	1	
ELISA	Not Reported	Positive	2	
ELISA	Chorus	Positive	1	
ELISA	Euroimmun	Positive	1	
ELISA	Pishtaz Teb	Not Reported	2	
ELISA	Pishtaz Teb	Negative	4	
ELISA	Pishtaz Teb	Positive	54	
ELISA	Vircell	Positive	6	
Other	Diesse-Tec	Positive	1	
Other	Pishtaz Teb	Positive	1	

Copyright © 2022 IACLD. All rights reserved.