

برگشت (/IACLD/Panel/Labs/Answers/29630)

دوره چهل و دوم تابستان 1401

## S2 Results

## نمونه مجهول ToRCH

کد نمونه 1161	EQAP-18136	آزمایشگاه پاتوبیولوژی و ژنتیک رصد
---------------	------------	-----------------------------------

Anti-CMV (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

Anti-Rubella (IgG)

Anti-Toxoplasma (IgG)

## Anti-HSV1 (Anti-Herpes Type1 IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است.

- ثبت نام کنندگان 215
- شرکت کنندگان 195
- کل درصد مشارکت 90.70%
- پاسخهای دریافتی 78
- درصد پاسخدهی 40.00%

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	12	12	0	0	%100.00	%0.00	%0.00
ECL	5	5	0	0	%100.00	%0.00	%0.00
ELFA	2	2	0	0	%100.00	%0.00	%0.00
ELISA	52	51	0	1	%98.08	%0.00	%1.92
Immunochromatography	1	1	0	0	%100.00	%0.00	%0.00
Other	6	6	0	0	%100.00	%0.00	%0.00

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
CL	Autobio Diagnostic	Positive	1	
CL	Dia Sorin	Positive	7	
CL	Liaison	Positive	3	
ECL	Abbott	Positive	1	

Method	Kit	Result	Count	Your Answer
ECL	Cobas	Positive	1	
ECL	Pishtaz Teb	Positive	1	
ECL	Roche	Positive	2	
ELFA	Biomerieux	Not Reported	1	
ELFA	Diesse-Tec	Positive	1	
ELFA	Orgentec	Positive	1	
ELISA	Not Reported	Not Reported	1	
ELISA	Not Reported	Positive	3	
ELISA	Architect	Positive	1	
ELISA	Autobio Diagnostic	Positive	1	
ELISA	Chorus	Positive	6	
ELISA	Euroimmun	Positive	2	
ELISA	ImmunoLab	Positive	1	
ELISA	Lion	Negative	1	
ELISA	Monobind	Not Reported	1	
ELISA	Orgentec	Positive	4	
ELISA	Pishtaz Teb	Not Reported	1	
ELISA	Pishtaz Teb	Positive	9	←
ELISA	Vircell	Positive	24	
Immunochromatography	Chorus	Positive	1	
Other	Chorus	Positive	2	
Other	Diesse-Tec	Positive	2	
Other	Orgentec	Positive	2	

Copyright © 2022 IACLD. All rights reserved.