

برگشت (/IACLD/Panel/Labs/Answers/32165)

دوره چهل و چهارم زمستان 1401

## S2 Results

## نمونه مجهول ToRCH

کد نمونه 1064	EQAP-18136	آزمایشگاه پاتوبیولوژی و ژنتیک رصد
---------------	------------	-----------------------------------

Anti-CMV (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

Anti-Rubella (IgG)

Anti-Toxoplasma (IgG)

## Anti-Toxoplasma (IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است.

- ثبت نام کنندگان 207
- شرکت کنندگان 185
- کل درصد مشارکت 89.37 %
- پاسخهای دریافتی 173
- درصد پاسخدهی 93.51 %

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	53	43	2	8	%81.13	%3.77	%15.09
ECL	29	28	0	1	%96.55	%0.00	%3.45
ELFA	22	12	10	0	%54.55	%45.45	%0.00
ELISA	67	59	5	3	%88.06	%7.46	%4.48
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00
Other	2	0	1	1	%0.00	%50.00	%50.00

Method	Kit	Result	Count	Your Answer
CL	Abbott	Negative	1	
CL	Abbott	Positive	8	
CL	Architect	Negative	1	
CL	Architect	Positive	2	
CL	Autobio Diagnostic	Positive	3	

Method	Kit	Result	Count	Your Answer
CL	Dia Sorin	Negative	3	
CL	Dia Sorin	Positive	20	←
CL	Liaison	Not Reported	1	
CL	Liaison	Indeterminate	1	
CL	Liaison	Negative	2	
CL	Liaison	Positive	3	
CL	Maglumi	Positive	2	
CL	Ortho	Indeterminate	1	
CL	Siemens	Negative	1	
CL	Siemens	Positive	5	
ECL	Abbott	Positive	2	
ECL	Architect	Positive	3	
ECL	Cobas	Positive	1	
ECL	Dia Sorin	Positive	4	
ECL	Liaison	Negative	1	
ECL	Liaison	Positive	1	
ECL	Roche	Positive	17	
ELFA	Biomerieux	Indeterminate	10	
ELFA	Biomerieux	Positive	12	
ELISA	Not Reported	Positive	2	
ELISA	Chorus	Indeterminate	4	
ELISA	Chorus	Negative	1	
ELISA	Chorus	Positive	2	
ELISA	Euroimmun	Indeterminate	1	
ELISA	Pishtaz Teb	Negative	2	
ELISA	Pishtaz Teb	Positive	49	
ELISA	Vircell	Positive	6	
Other	Chorus	Indeterminate	1	
Other	Chorus	Negative	1	